

MANUFACTURER'S DECLARATION: MEDICAL DEVICES REGULATION 2017/745

As this component is not a complete "Medical Device" under the definition provided in the Regulation but a part with no function on its own, a Declaration of Conformity or CE mark is not appropriate until the component is incorporated. The following Declaration is issued in order to aid the installation of the component below into a finished appliance, for use by, or sale to, an enduser.

The following declares the status of these products with regard to Union Regulation 2017/745:

Product Description	Product Name	Part Numbers
R-net Joystick	R-net JSM-LED	D51108, D51122,
Module LED		D51315, D51316, D51623

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation to the extent possible as a component to be integrated into a complete product.

APPLICABLE STANDARDS

The following standards (and those called by them) have been used in order to assess a presumption of conformity with the essential requirements of the above regulation as far as the component allows:

Medical devices — Application of risk management to medical devices
Medical device software — Software life-cycle processes
Biological evaluation of medical devices - Part 1: Evaluation and testing within
a risk management process
Electrically powered wheelchairs, scooters and their chargers. Requirements
and test methods
Wheelchairs – Part 9: Climatic tests for electric wheelchairs
Wheelchairs Part 14: Power and control systems for electrically powered
wheelchairs and scooters Requirements and test methods
Wheelchairs Part 21: Requirements and test methods for electromagnetic
compatibility of electrically powered wheelchairs and scooters, and battery
chargers

Nigel D Mills 14 June 2023

Nigel Mills, Senior Manager, Engineering

Signed at, for and on behalf of:

Penny & Giles Controls Ltd.,

15 Enterprise Way, Aviation Park West, Christchurch, Dorset, UK BH23 6HH